



STATE OF NEW YORK - DEPARTMENT OF LABOR
DIVISION OF SAFETY AND HEALTH

RADIOACTIVE MATERIALS LICENSE

DL 05-239

Page 1 of 4 Page(s)

PURSUANT TO THE LABOR LAW AND INDUSTRIAL CODE RULE 38, AND IN RELIANCE ON STATEMENTS AND REPRESENTATIONS HERETOFORE MADE BY THE LICENSEE DESIGNATED BELOW, A LICENSE IS HEREBY ISSUED AUTHORIZING SUCH LICENSEE TO RECEIVE, POSSESS, USE AND TRANSFER RADIOACTIVE MATERIALS(S) DESIGNATED BELOW; AND TO USE SUCH RADIOACTIVE MATERIALS FOR THE PURPOSE(S) AND AT THE PLACE(S) DESIGNATED BELOW. THIS LICENSE IS SUBJECT TO ALL APPLICABLE RULES, REGULATIONS, AND ORDERS NOW OR HEREAFTER IN EFFECT OF ALL APPROPRIATE REGULATORY AGENCIES AND TO ANY CONDITIONS SPECIFIED BELOW.

| | | | |
|--|----------------------------------|---|--|
| 1. NAME OF LICENSEE Testing Mechanics Corp. | | FEIN: 11-2957640 PHONE:(516) 221-3800 | 3. LICENSE NUMBER 2672-3934 |
| 2. ADDRESS OF LICENSEE 3770 Merrick Road Seaford, New York 11783 | | | 4. EXPIRATION DATE August 31, 2008 |
| | | | 5a. REFERENCE NO. <u>3</u> b. AMENDMENT NO. <u>1</u> |
| 6. Radioactive materials (element in mass number) | 7. Chemical and/or physical form | 8. Maximum quantity licensee may possess at any one time | |
| A. Cesium 137 | A. Sealed Sources | A. See Condition 9 | |
| B. Americium 241: Beryllium | B. Sealed Sources | B. See Condition 9 | |
| C. Cobalt 57 | C. Sealed Source | C. See Condition 9 | |
| D. Cadmium 109 | D. Sealed Source | D. See Condition 9 | |

9. Authorized use:

Conditions 6.A. and 6.B.

- A. The licensee is authorized to use any sealed source, or associated portable moisture/density gauge that has been manufactured and distributed in accordance with a specific license issued by an Agreement State or the United States Nuclear Regulatory Commission. Combinations of sources and devices must be compatible for use as stated in a Sealed Source and Device Registration Certificate (i.e. stated in the registration certificate for the source or device).
- B. No single source may exceed the maximum activity specified for that nuclide in the Sealed Source and Device Registration Certificate for any device in which the source is to be used.

Conditions 6.C. and 6.D.

- A. The licensee is authorized to use any sealed source, or associated portable x-ray fluorescence device that has been manufactured and distributed in accordance with a specific license issued by an Agreement State or the United States Nuclear Regulatory Commission. Combinations of sources and devices must be compatible for use as stated in a Sealed Source and Device Registration Certificate (i.e. stated in the registration certificate for the source or device).