



NEW YORK STATE DEPARTMENT OF HEALTH  
RADIOACTIVE MATERIALS LICENSE

Pursuant to the Public Health Law, Part 16 of the New York State Sanitary Code, Industrial Code Rule 38, and in reliance on statements and representations heretofore made by the licensee designated below, a license is hereby issued authorizing radioactive material(s) for the purpose(s), and at the place(s) designated below. The license is subject to all applicable rules, regulations, and orders now or hereafter in effect of all appropriate regulatory agencies and to any conditions specified below.

<p>1. NAME OF LICENSEE</p> <p style="text-align: right;">FEIN 11-2957640</p> <p><b>Testing Mechanics Corporation</b></p> <p style="text-align: right;">Phone (516) 221-2333</p>	<p>3. LICENSE NUMBER</p> <p style="text-align: center;">C2672</p> <hr/> <p>4. EXPIRATION DATE</p> <p style="text-align: center;"><b>February 14, 2027</b></p>
<p>2. ADDRESS OF LICENSEE</p> <p>3770 Merrick Road Seaford, New York 11783</p>	<p>5a. REFERENCE                      b. AMENDMENT NO.</p> <p style="text-align: center;"><b>DH 11-713                      4</b></p>

6. Radioactive Materials (elements in mass number)	7. Chemical and/or physical form	8. Maximum quantity licensee may possess at any one time
A. Cesium 137	A. Sealed sources	A. Not to exceed 200 millicuries total
B. Americium 241/ Beryllium	B. Sealed sources	B. Not to exceed 800 millicuries total
C. Cobalt 57	C. Sealed sources	C. 24 millicuries

9. Authorized use.  
Conditions 6.A. and 6.B.:
- A. The licensee is authorized to use any sealed source or associated portable moisture/density gauge which has been manufactured and distributed in accordance with a specific license issued by an Agreement State or the United States Nuclear Regulatory Commission. Combinations of sources and devices must be compatible for use as stated in a Sealed Source and Device Registration Certificate (i.e., stated in the registration certificate for the source or device).
  - B. No single source may exceed the maximum activity specified for that nuclide in the Sealed Source and Device Registration Certificate for any device in which the source is to be used.